E-197 PULSERIDER FOR ANEURYSM RE-TREATMENT: COMBINED SUB-ANALYSIS OF THE ANSWER AND AWARD STUDIES

¹A Spiotta*, ²D Lopes, ³J Mocco, ⁴A Siddiqui, ⁵S Tateshima, ⁶H Nahser. ¹Department of Neurosurgery, University Hospital, Medical University of South Carolina, Charleston, SC; ²Brain and Spine Institute, Advocate Aurora Health, RUSH University, Park Ridge, IL; ³Department of Neurosurgery, Mount Sinai Hospital, New York, NY; ⁴Departments of Neurosurgery and Radiology, Jacobs School of Medicine and Biomedical Sciences, University at Buffalo, New York, Buffalo, NY; ⁵Department of Radiological Sciences, University of California, Los Angeles, Diagnostic and Interventional Radiology Diagnostic Radiology, Los Angeles, CA; ⁶The Walton Centre NHS Foundation Trust – Emeritus, Liverpool, UK

10.1136/neurintsurg-2020-SNIS.228

Introduction PulseRider is a neck bridging device designed to aid in coiling of wide-neck intracranial aneurysms. When in place, it provides a ‘floor’ to the coil mass while allowing for easy access to the aneurysm sac, making it potentially suited for retreatment cases with limited working space.

Methods Procedural success and outcomes of retreatment cases included in the AWARD (Safety Study of the PulseRider in Patients Undergoing Treatment for Bifurcation Intracranial Aneurysms) and ANSWER (Adjunctive Neurovascular Support for Wide-neck aneurysm Embolization and Reconstruction) studies are presented.

Results Eleven PulseRider cases (10 female, mean age 50.9 years), 6 in AWARD and 5 in ANSWER, were retreatments of previously coiled aneurysms. Eight of the patients had 1 previous procedure, 2 had 2 previous treatments, and 1 had 3, including a failed attempt at Y-stenting. Mean aneurysm neck size was 4.98 mm (range 3.0–11.6 mm). Locations included basilar (7), carotid terminus (2) and MCA (2). PulseRider was successfully implanted with the ability to retain coils within the aneurysm in all cases. Immediately post-procedure, core lab adjudicated adequate occlusion rate was 91% (4/11 [36%] RR-I, 6/11 [55%] RR-II and 1/11 [9%] RR-III). At the one year follow up, the adjudicated rate of adequate aneurysm occlusion was 80% (6/10 [60%] RR-I, 2/10 [20%] RR-II, and 2/10 [20%] RR-III). There were no neurological deaths or major ipsilateral stroke (primary safety endpoint in the ANSWER study) and neurological decline was noted in one patient (the case of 3 previous treatments).

Conclusion Retreatment of previously coiled wide-neck aneurysms with the PulseRider device resulted in 100% procedural success and 80% complete or near complete occlusion at 1 year with an acceptable safety profile.

Disclosures A. Spiotta: 1; C; Microvention. 2; C; Penumbra, Minnetronix, Cerenovus. 6; C; Cerenovus, Penumbra, Microvention. D. Lopes: None. J. Mocco: 2; C; Rebound Medical, LLC, Cerebrotech Medical Systems, Inc., Synchron, Inc., EndoStream Medical, Ltd.. 4; C; Apama Medical, Inc., NeuroTechnology Investors, LLC., Neurvana Medical, LLC, Cerebrotech Medical Systems, Inc., Synchron, Inc., EndoStream Medical, Ltd., Stroke Project, Inc. A. Siddiqui: 2; C; Amnis Therapeutics, Boston Scientific, Canon Medical Systems USA, Cerebrotech Medical Systems, Claret Medical, Corindus, Endstream Medical, Guidepoint Global Consulting, Imperative Care, Integra, Rapid Medical, Rebound Therapeutics Corp., Silk Road Medical, StimMed, Stryker, Three Rivers Medical, VasSol, W.L. Gore & Associates, Cerenovus, Medtronic, MicroVenting, Penumbra, Medical University of South Carolina, National PI/Steering Committee for POSITIVE Trial, Northwest University, DSMB Chair for HEAT Trial. 4; C;

Amnis Therapeutics, Apama Medical, BlinkTBI, Buffalo Technology Partners, Cardinal Health, Cerebrotech Medical Systems, Claret Medical, Cognition Medical, Endostream Medical, Ltd, Imperative Care, International Medical Distribution Partners, Rebound Therapeutics Corp, Silk Road Medical, StimMed, Synchron, Three Rivers Medical, Viseon. S. Tateshima: 2; C; Cerenovus, Medtronic, Neurovasc, Stryker. 4; C; Neurovasc. H. Nahser: None.

E-198 FACIAL PAIN CAUSED BY A CAVERNOUS SINUS ANEURYSM TREATED WITH ENDOVASCULAR FLOW DIVERSION: CASE REPORT AND REVIEW OF THE LITERATURE

A Larson*, T Mehta, A Grande. Neurosurgery, University of Minnesota, Minneapolis, MN

10.1136/neurintsurg-2020-SNIS.229

Flow diversion is a commonly-used endovascular modality to treat intracavernous aneurysms. Cases of intracavernous aneurysms causing facial pain are rare, and flow diversion may represent a viable treatment option in order to alleviate pain caused by these aneurysms. We present the case of a 53-year-old woman who presented with a two-year history of persistent, left-sided and facial pain and numbness located in the left V1-V2 distribution of the Trigeminal nerve. A brain MRI was obtained which showed a left internal carotid artery aneurysm projecting infero-laterally within the cavernous sinus and possibly compressing left V1 and V2 roots of the Trigeminal nerve. This was further investigated and confirmed with digital subtraction angiography. To reduce mass effect, the aneurysm was treated with flow-diversion rather than coil-embolization. Her pain and numbness gradually improved with near-complete resolution at 6 months post-treatment. 6-month follow-up angiography demonstrated minimal residual filling of the aneurysm suggesting a reduction in size of the aneurysm sac. This case, in addition to prior reports which we review here, demonstrate the ability of flow diversion to treat facial pain secondary to intracavernous aneurysms.

Disclosures A. Larson: None. T. Mehta: None. A. Grande: None.

E-199 PRELIMINARY EXPERIENCE WITH SURPASS EVOLVE FLOWDIVERTER DEVICE: CLINICAL AND TECHNICAL NOTE

¹K Alpay*, ¹R Rautio, ²J Numminen, ¹M Sinisalo. ¹Department of Radiology, Turku University Hospital, Turku, Finland; ²Department of Radiology, Helsinki University Hospital, Helsinki, Finland

10.1136/neurintsurg-2020-SNIS.230

Objective The aim of the study was to evaluate safety and efficacy of the new generation flow diverter device, Surpass Evolve (SE).

Methods In two Finnish early-user centers, twenty-seven patients with ruptured or unruptured intracranial aneurysms were treated with SEs. All procedure-related events were recorded into the local database for further analysis. We report here our results concerning the clinical and radiological

Abstract E-199 Table 1 Patients, aneurysms and device characteristics

	Age	Location	Type of Aneurysm	Size (maximal diameter)	UA/RA	Size of Surpass Evolve	Intraprocedural Complication	Post Procedural Complication	Immediate Radiological Outcome (OKM)
1	58	Right ICA	Saccular	10 mm	UA	4x15 mm	No	No	A3
2	61	Left ICA	Saccular	4 mm	UA	4.5x20 mm	No	No	A1
3	56	Left ICA	Saccular	5 mm	UA	5x20 mm	No	No	B3
4	62	Right ICA	Saccular	4 mm	UA	4.5x15 mm	No	No	C2
5	52	Left ICA	Saccular	4 mm	UA	4.5x17 mm	No	No	C3
6	52	Left MCA trunk	Saccular	4 mm	UA	3.25x12 mm	No	No	B3
7	66	Basilar trunk	Saccular	6 mm	RA	4x12 mm + 4x12 mm	No	No	C2
8	57	Left MCA trunk	Saccular	3 mm	UA	2.5x12 mm	No	No	C3
9	72	Right ICA	Saccular	10 mm	UA	4.5x25 mm	No	No	B2
10	59	Left Vertebral	Saccular	3 mm	RA	3.25x17 mm	No	No	B2
11	30	Left ICA	Saccular	3 mm / 3 mm	UA	4.5x17 mm	No	No	B3
12	64	Right ICA	Saccular	6 mm	UA	4x12 mm	No	No	C3
13	66	Right ICA	Saccular	4 mm	UA	4.5x17 mm	No	No	A3
14	61	Left ICA	Saccular	24 mm	UA	5x20 mm	No	No	B3
15	52	Right ICA	Saccular	15 mm	UA	5x20 mm	No	No	A3
16	48	Left ICA	Saccular	4 mm	UA	5x20 mm	No	No	A1
17	48	Right ICA	Saccular	6 mm	UA	5x20 mm	No	No	A3
18	60	Left ICA	Saccular	18 mm	UA	4x20 mm	No	No	A3
19	41	Left ICA	Saccular	7 mm	UA	4.5x15 mm	No	No	A3
20	32	Left ICA	Saccular	10 mm	UA	4x17 mm	No	No	B3
21	32	Left ICA	Saccular	4 mm	UA	4.5x17	No	No	A3
22	70	Right ICA	Saccular	14 mm	UA	4.5x15 mm + 4.5x12 mm	No	No	A3
23	42	Right ICA	Saccular	18 mm	UA	4.5x25 mm	No	No	B3
24	49	Right ICA	Saccular	7 mm	UA	5x15 mm	No	No	A2
25	68	Right ICA	Blister	3 mm	UA	4.5x12 mm	No	No	A1
26	55	Right Vertebral	Fusiform	6 mm	RA	3.25x20 mm	No	No	C3
27	52	Vertebrobasilar	Fusiform	28 mm	RA	5x40 mm	No	No	B3

RA, ruptured aneurysm; UA, unruptured aneurysm; OKM, O'Kelly Marotta Classification

results from intracranial aneurysms treated with this new generation flow diverter device, SE.

Results Twenty-eight aneurysms (86% unruptured) were treated with twenty-nine SE devices. All procedures were technically successful. In one aneurysm (case #26), a coil was used to secure the thrombosis. At discharge, all patients had a modified Rankin score of 0–2. No intraprocedural or thromboembolic complications were encountered. One subarachnoid hemorrhage (SAH) occurred one week after the treatment (case #22), and this patient died due to SAH. The median follow-up time was two months (range; 0–6 months). The radiological follow-up data was available for nine patients (33%); two aneurysms were occluded during the follow-up period.

Conclusions Our initial experience with the new generation SE indicates that the device is safe to use. The new generation SE provides reliable deployment and lower foreshortening. Mid-term and long-term follow-up examinations will be needed to provide data on longer term safety outcomes.

Disclosures K. Alpay: None. R. Rautio: 2; C; Microvention, Stryker, Medtronic. J. Numminen: None. M. Sinisalo: None.

E-200

GENDER AND RACIAL DISPARITY IN OUTCOME OF ANEURYSMAL SUBARACHNOID HEMORRHAGE IN THE UNITED STATES

S Majidi*, A Mehta, B Yim, K Yaeger, R De Leacy, J Mocco, J Fifi. *Neurosurgery, Mount Sinai Hospital, New York, NY*

10.1136/neurintsurg-2020-SNIS.231

Background Aneurysmal subarachnoid hemorrhage (aSAH) is associated with a high rate of morbidity and mortality. We aimed to assess the prognostic impact of gender and race in these patients.

Methods The Nationwide Inpatient Sample (NIS 2006–2016) was utilized to identify patients presenting with aSAH as primary diagnosis. Patient age, sex, race/ethnicity, insurance status, socioeconomic status, comorbidities, type of the hospital and treatment modality used for aneurysm repair were extracted. The previously validated NIS-SAH Severity Scale was used to estimate the clinical severity. Discharge destination and in-hospital mortality were used as outcome measures. The impact of race and gender on clinical outcome was analyzed using multivariate regression models.

Results A total of 19,512 patients with aSAH were identified. Mean age ($\pm$ SD) was 54 ($\pm$ 0.1). Sixty seven percent of the patients were female, 49% white and 13% black. Female patients were 4 years older (55 vs 51, $p<0.0001$), and had higher numbers of comorbidities. There was no difference in the NIS-SAH Severity Scale score between the two genders. Women had significantly lower odds of good clinical outcome (defined as discharge to home or acute rehabilitation facility) after adjusting for confounding factors [OR, (95% CI); p -value: 0.88, (0.81–0.96); $p<0.0001$]. In regard to race, Hispanic patients [OR, (95% CI); p -value: 1.49, (1.29–1.72); $p<0.0001$] and Asians [OR, (95% CI); p -value: 1.33, (1.07–1.64); $p<0.0001$] had higher odds of good clinical outcome after adjusting for age, gender, SAH severity, comorbidities and treatment modality. Lower odds of mortality was observed in black [OR, (95% CI); p -value: 0.55, (0.46–0.66); $p<0.0001$] and Hispanic patients [OR, (95% CI); p -value: 0.72, (0.59–0.88); $p=0.001$] compared to the white patients.

Conclusions In this nationally representative study, women and white patients had disproportionately higher likelihood of poor clinical outcome. Lower rate of mortality is seen among black and Hispanics.

Disclosures S. Majidi: None. A. Mehta: None. B. Yim: None. K. Yaeger: None. R. De Leacy: None. J. Mocco: None. J. Fifi: None.