Abstract E-036 Figure 1

Introduction Flow diverters (FDs) are sized to the recipient vessel during pre-treatment planning. However, sizing can be challenging because of large changes in vessel curvature and diameter. Further, FDs can elongate by more than 50% of their labeled length after deployment, which complicates sizing. Significant complications can result from over- or under-sized devices. Here we present a finite element (FE) modelling approach for evaluating FD size and compare that approach to clinical deployments to determine its accuracy in predicting device length, diameter, and apposition.

Methods Ten patient cases treated with the pipeline embolization device were acquired from two hospitals. Pre- and post-treatment CT image data were segmented then reconstructed to form computational models of the devices and vessels. A library of pipeline FE models, which was previously validated against physical devices, was used to simulate deployment of the same devices into the pre-treatment vessels. The pipeline models were first navigated via a virtual microcatheter to the landing zones observed in the post-treatment vessels. A “push-pull” algorithm was then used to simulate device unsheathing. Three post-deployment metrics were compared: device apposition to the vessel wall along the vessel centerline, device diameter along the stent centerline, and device length.

Results Simulated and clinical deployments showed good agreement both qualitatively and quantitatively (Figure 1). Simulations captured regions where the device poorly apposed to the vessel wall and poorly covered the aneurysm neck. Mean errors between simulated and clinical deployments (as a percentage of the clinical value) were less than 5% for device length and 9% for device apposition and diameter.

Conclusion FE simulations captured post-deployment FD shape and apposition. Less than a 9% mean error was found between simulated and clinical deployment metrics. These results provide additional support for the use of FE for evaluating device size during pre-treatment planning.

Disclosures B. Chong: 4; C; Endovantage, LLC. H. Babiker: 4; C; Endovantage, LLC. Y. Kalani: None. C. Baccin: None. M. Mortensen: 5; C; Endovantage, LLC. M. Levitt: None. C. McDougall: None. D. Frakes: 4; C; Endovantage, LLC. F. Albuquerque: 4; C; Endovantage, LLC.

Introduction Facial AVMs are associated both with risk of hemorrhage and considerable psychosocial as well as cosmetic distress. Head and neck AVMs are the second most common outside of the brain, but despite being a richly-vascularized area, giant and diffuse lesions of the magnitude reported are rare and typically evolve from iatrogenesis.

Case presentation A 32 year-old male presented with a pulsatile facial mass which had treated unsuccessfully at numerous outside intuitions. Imaging revealed a diffuse facial AVM with extensive bilateral supply, and a ligated left external carotid artery. Though incurable, endovascular treatment was required frequently for hemorrhagic events and palliation were delivered via the contralateral supply due ligation of a direct route to the nidus.

Discussion We report a particularly complicated case of giant facial AVM which initially imparted significant psychosocial morbidity, but as a result of proximal ligation and improper embolization grew into a tremendous lesion the size and extent of which has not been reported previously. It remains absolutely essential that no attempt at palliation or cure be made that cannot directly target the nidus of an AVM, and this ever the more crucial in rich-supplied areas such as the face. Endovascular embolizations remains the first and often only option during when emergent therapy.

Abstract E-037 Figure 1

Disclosures A. Dmytriw: None. J. Song: None. S. Power: None. R. Agid: None.

E-037 DIFFUSE FACIAL ARTERIOVENOUS MALFORMATION WITH LIGATED IPSILATERAL ARTERIAL SUPPLY

¹A Dmytriw, ²J Song, ³S Power, ³R Agid. ¹Department of Medical Imaging, University of Toronto, Toronto, ON, Canada; ²Department of Otolaryngology, University of Toronto, Toronto, ON, Canada; ³Department of Medical Imaging, Toronto Western Hospital, Toronto, ON, Canada

10.1136/neurintsurg-2016-012589.109

E-038 3D CONE BEAM COMPUTERIZED TOMOGRAPHY (3D-CBCT) GUIDED SACROILIAC (SI) JOINT INJECTION: A REAL-TIME, INTERACTIVE, ACCURATE, FAST AND REDUCED RADIATION EXPOSURE TECHNIQUE

¹S Lee, ¹S Ali, ²J Mok. ¹Radiology, University of Chicago, Chicago, IL; ²Orthopedic Surgery, University of Chicago, Chicago, IL

10.1136/neurintsurg-2016-012589.110

Abstract E-038 Figure 1

Purpose To ensure effective therapeutic management of sacroiliac (SI) joint pain, precise needle placement during diagnostic SI joint injection is essential. The reported accuracy of diagnostic SI pain assessment using single or dual blocks is highly variable, with false positive rate of 10–30%. CT guidance for sacroiliac joint injection has been the preferred imaging technique for needle guidance due to the complex 3 dimensional anatomy of the SI joint, frequent encounter of covering osteophytes and occasionally encountered extremely narrowed SI joint. However, the technique has been challenged due to relatively prolonged procedure time and risk of increased radiation exposure. We evaluated the technical feasibility of the 3D-CBCT guided diagnostic SI joint injection technique.

Materials and methods Four patients (5 SI joints) with clinically suspected SI joint pain syndrome were enrolled. The procedure was performed under local anesthesia in a bi-plane neuroangiography suite (AlluraXper®, Philips, Netherlands), but only AP plane was used for the procedure. Using the integrated image guidance system (XperGuide®, Philips, Netherlands), the target and needle trajectory were selected based on 3D-CBCT images. Subsequently, the selected images were fused with real-time fluoroscopy for the needle guidance. Post needle placement and post contrast injection 3D-CBCTs were also obtained before the injection of local anesthetics (approximately 0.5–1.0 ml) into the target SI joint.

Results The needle placement and diagnostic injection for the target SI joint were successful in all 4 cases (5 SI joints). Median fluoroscopy time per SI joint was 5.2 min (range: 4.6 min – 6.7 min) and total DAP was 87,445.2 mGycm² (range: 32,864 mGycm² – 199,303 mGycm²). Pre-procedure mean pain level was 7.5/10 and immediate post-procedure mean pain level was 0/10.

Conclusion 3D-CBCT guided diagnostic SI joint injection is technically feasible, effective, reasonably fast and can be performed with relatively low radiation exposure. In addition, the real-time, interactive nature of the guidance system may provide more confidence to the operator.

Disclosures S. Lee: None. S. Ali: None. J. Mok: None.

E-039 COMPARISON OF MEDICAL THERAPY VS. INTRA-ARTERIAL THERAPY FOR ACUTE VERTEBROBASILAR STROKE

¹S Raymond, ²T Leslie-Mazwi, ²N Rost, ¹P Schaefer, ¹J Hirsch, ¹R Gonzalez, ¹J Rabinov. ¹Radiology, Massachusetts General Hospital, Boston, MA; ²Neurology, Massachusetts General Hospital, Boston, MA

10.1136/neurintsurg-2016-012589.111

Introduction/purpose We present a consecutive retrospective cohort of patients treated at an academic stroke center for

acute ischemic stroke from large vessel occlusion of the verte-brobasilar circulation. We describe prognostic factors and outcomes for patients receiving medical and/or intra-arterial therapy (IAT) over a 12 year period.

Materials and methods We treated 87 patients (medical therapy N = 44 and IAT = 43) between 2004–2015 for acute posterior circulation infarct with verte-brobasilar occlusion at our institution. Patients were dichotomized by modified Rankin Score (mRS) to good outcome defined as mRS 0–2 or poor outcome defined as mRS 3–6. Multiple characteristics were compared including age, gender, NIHSS, infarct volume, posterior ASPECTS, thrombus location, hconversion and treatment modality (medical therapy or IAT). Categorical variables were compared using Chi-square statistic, and continuous variables were compared using t-test and Mann-Whitney statistics. The study was conducted with IRB approval.

Results Major etiologies of acute basilar occlusion included atrial fibrillation, PFO, vertebral artery dissection with distal embolus, and intracranial atherosclerosis (ICAD). Poor outcomes were associated with higher NIHSS, lower ASPECTS and extensive brainstem involvement on DWI. IAT patients presented with more severe deficit (NIHSS 21) compared to patients treated medically (NIHSS 14). Rates of good outcomes (mRS 0–2) were equivalent between the IAT (21/43, 49%) and medically treated cohort (21/44, 48%). When ICAD patients were excluded from analysis, results were similar with 50% of the IAT patients (18/36) and 43% of medically treated patients (15/35) having a good outcome. This time period spans major changes in endovascular techniques ranging from urokinase and wire manipulation to current devices. ADAPT or stentrievors were only used in ten cases.

Conclusion Compared to medical trials in 2015, which demonstrate clinical benefit for IAT over medical therapy for acute ischemic stroke of the anterior circulation, the posterior circulation is less well studied. Important prognostic factors for stroke due to large vessel occlusion of posterior circulation include posterior ASPECTS score and infarct volume in the brainstem. In this small retrospective study, patients treated with IAT had greater stroke severity than the medical cohort, yet they had similar rates of good functional outcome. Parameters of this retrospective analysis may be used to improve triage and treatment of patients with acute verte-brobasilar occlusion for thrombectomy using current technology.

REFERENCES

- 1 Singer OC, Berkefeld L, Nolte CH, et al. Mechanical recanalization in basilar artery occlusion: The ENDOSTROKE Study. *Ann Neurol* 2015 Mar;77(3):415–24.
- 2 Strbian D, Sairanen T, Silvennoinen H, et al. Thrombolysis of basilar artery occlusion: impact of baseline ischemia and time. *Ann Neurol* 2013 Jun;73(6):688–94