

	Aneurysm location	Type of aneurysm	Stent	Type of Stenting	Reason to remove	Alligator	Snare	Solitaire	Lasso & Solitaire	Snare & Wire	Allover Result	mRS at discharge
1	Acom	Unruptured	Leo	Regular	Unexpansion	NA	Success	NA	NA	NA	Success	0
2	MCA	Recanalized	BB Leo	Regular	Thrombosis	Failure	Failure	Failure	Success	NA	Success	-
	Acom	Recanalized	BB Leo	Regular	Thrombosis	Failure	Failure	Failure	Failure	NA	Failure	6
3	MCA	Unruptured	Silk	Regular	Displacement	NA	Success	NA	NA	NA	Success	0
4	Basilar	Unruptured	Fred	Regular	Thrombosis	NA	Failure	Failure	Failure	Success	Success	0
5	Pica	Unruptured	BB Leo	1/2 T stenting	Thrombosis	Failure	Success	NA	NA	NA	Success	1
6	Pcom	Recanalized	Leo	T stenting	Hemorrhage	Failure	Success	NA	NA	NA	Success	0
			BB Leo	T stenting	Hemorrhage	Failure	Success	NA	NA	NA	Success	0
7	Vert V4	Unruptured	Derivo	Regular	Thrombosis	Failure	Failure	Failure	Failure	NA	Failure	3
8	Paraophthalmic	Unruptured	Leo	Regular	Displacement	NA	NA	Success	NA	NA	Success	0
9	Pcom	Unruptured	Leo	Regular	Unexpansion	Failure	Failure	Failure	Failure	NA	Failure	0
10	Pcom	Recanalized	Silk	Kissing stents	Displacement	NA	Success	Success	NA	NA	Success	0
			Silk	Kissing stents	Displacement	NA	Success	Success	NA	NA	Success	0
11	Carotid T	Unruptured	Leo	T stenting	Thrombosis	Failure	Success	NA	NA	NA	Success	0
			BB Leo	T stenting	Thrombosis	NA	Success	NA	NA	NA	Success	0
12	Pcom	Recanalized	Leo	Regular	Unexpansion	NA	Success	NA	NA	NA	Success	0
13	Basilar	Unruptured	BB Leo	1/2 T stenting	Thrombosis	NA	NA	Failure	Success	NA	Success	0
14	Basilar	Unruptured	Acclion	Regular	Displacement	NA	NA	Success	NA	NA	Success	0
			BB Leo	Regular	Displacement	NA	NA	NA	Failure	Success	Success	0
15	Siphon	Unruptured	Leo	Regular	Unexpansion	NA	NA	NA	NA	Success	Success	0

16	Pcom	Unruptured	Leo	T stenting	Thrombosis	Failure	Success	NA	NA	NA	Success	-
			BB Leo	T stenting	Thrombosis	Failure	Failure	Failure	NA	NA	Failure	3
17	Paraophthalmic	Unruptured	Lvis	Regular	Unexpansion	NA	NA	Success	NA	NA	Success	0
18	Acom	Unruptured	BB Leo	Regular	Thrombosis	Failure	Failure	Failure	Success	NA	Success	0
19	Cavernous	Unruptured	Fred	Telescopic	Unexpansion	NA	Success	NA	NA	NA	Success	0
			Fred	Telescopic	Unexpansion	NA	Success	NA	NA	NA	Success	0
20	Acom	Recanalized	Lvis	Regular	Coil migration	NA	Success	NA	NA	NA	Success	0
21	MCA	Unruptured	BB Leo	T stenting	Thrombosis	NA	NA	NA	Success	NA	Success	0
22	Acom	Unruptured	BB Leo	Regular	Thrombosis	NA	Failure	Failure	Success	NA	Success	0
23	Cavernous	Unruptured	Leo	Regular	Thrombosis	NA	NA	Success	NA	NA	Success	0
24	Carotid T	Unruptured	Lvis	Regular	Thrombosis	NA	Success	NA	NA	NA	Success	0
25	MCA	Unruptured	BB Leo	Regular	Thrombosis	NA	Failure	Failure	Success	NA	Success	0
26	MCA	Recanalized	Lvis	Regular	Unexpansion	Failure	Success	NA	NA	NA	Success	0
27	MCA	Unruptured	BB Leo	1/2 T stenting	Thrombosis	NA	NA	NA	Success	NA	Success	0
28	Basilar	Unruptured	Leo	T stenting	Thrombosis	NA	NA	NA	Success	NA	Success	0
			BB Leo	T stenting	Thrombosis	NA	NA	NA	Success	NA	Success	0
29	Basilar	Unruptured	BB Leo	Regular	Unexpansion	NA	Success	NA	NA	NA	Success	0
30	Basilar	Recanalized	Pipeline	Telescopic	Occlusion of side branch	Failure	Failure	Failure	Success	NA	Success	0
			Pipeline	Telescopic	Occlusion of side branch	Failure	Failure	Failure	Success	NA	Success	0

			Pipeline	Telescopic	Occlusion of side branch	Failure	Failure	Failure	Success	NA	Success	0
31	Basilar	Unruptured	Lvis	Kissing stents	Thrombosis	NA	Success	NA	NA	NA	Success	0
			Lvis	Kissing stents	Thrombosis	NA	Success	NA	NA	NA	Success	0
32	MCA	Unruptured	BB Leo	1/2 T stenting	Thrombosis	NA	Failure	Failure	Success	NA	Success	0
33	MCA	Recanalized	BB Leo	1/2 T stenting	Thrombosis	NA	NA	NA	Success	NA	Success	0
34	Acom	Unruptured	BB Leo	Regular	Misplacement	NA	NA	Failure	Success	NA	Success	0
35	Basilar	Unruptured	Lvis	Regular	Misplacement	NA	Failure	Failure	Success	NA	Success	0
36	MCA	Unruptured	BB Leo	Regular	Unexpansion	Failure	Failure	Failure	Failure	NA	Failure	0