









OPEN ACCESS

Original research

Evaluation of effectiveness and safety of the CorPath GRX robotic system in endovascular embolization procedures of cerebral aneurysms

Vitor Mendes Pereira ¹, Hal Rice,² Laetitia De Villiers,² Nader Sourour,³ Frédéric Clarencon,³ Julian Spears,¹ Alejandro Tomasello ⁴, David Hernandez,⁴ Nicole M Cancelliere ¹, Xiao Yu Eileen Liu,¹ Patrick Nicholson,⁵ Vincent Costalat,⁶ Gregory Gasco ⁶, Pasquale Mordasini,⁷ Jan Gralla,⁷ Mario Martínez-Galdámez ⁸, Jorge Galvan Fernandez,⁸ Monika Killer-Oberpfalzer ⁹, Raymond D Turner,¹⁰ Raphael Blanc,¹¹ Michel Piotin¹¹

For numbered affiliations see end of article.

Correspondence to

Dr Vitor Mendes Pereira, Department of Neurosurgery, Director of Endovascular Research and Innovation, St Michael's Hospital, Toronto, Canada; vitormpbr@hotmail.com

Received 7 February 2023
Accepted 7 May 2023

ABSTRACT

Background Robotic-assisted neurointervention was recently introduced, with implications that it could be used to treat neurovascular diseases.

Objective To evaluate the effectiveness and safety of the robotic-assisted platform CorPath GRX for treating cerebral aneurysms.

Methods This prospective, international, multicenter study enrolled patients with brain aneurysms that required endovascular coiling and/or stent-assisted coiling. The primary effectiveness endpoint was defined as successful completion of the robotic-assisted endovascular procedure without any unplanned conversion to manual treatment with guidewire or microcatheter navigation, embolization coil(s) or intracranial stent(s) deployment, or an inability to navigate vessel anatomy. The primary safety endpoint included intraprocedural and periprocedural events.

Results The study enrolled 117 patients (74.4% female) with mean age of 56.6 years from 10 international sites. Headache was the most common presenting symptom in 40/117 (34.2%) subjects. Internal carotid artery was the most common location (34/122, 27.9%), and the mean aneurysm height and neck width were 5.7±2.6 mm and 3.5±1.4 mm, respectively. The overall procedure time was 117.3±47.3 min with 59.4±32.6 min robotic procedure time. Primary effectiveness was achieved in 110/117 (94%) subjects with seven subjects requiring conversion to manual for procedure completion. Only four primary safety events were recorded with two intraprocedural aneurysm ruptures and two strokes. A Raymond-Roy Classification Scale score of 1 was achieved in 71/110 (64.5%) subjects, and all subjects were discharged with a modified Rankin Scale score of ≤2.

Conclusions This first-of-its-kind robotic-assisted neurovascular trial demonstrates the effectiveness and safety of the CorPath GRX System for endovascular embolization of cerebral aneurysm procedures.

Trial registration number NCT04236856

WHAT IS ALREADY KNOWN

⇒ Robotic-assisted neurovascular intervention was recently introduced, with only a handful of published cases.

WHAT THIS STUDY ADDS

⇒ This is the first-of-its-kind trial to assess the efficacy and safety of robotics use in neurovascular intervention. From this trial, we now know that robotic-assisted neurovascular intervention is an effective and safe method to embolize cerebral aneurysms.

HOW THIS STUDY MIGHT AFFECT RESEARCH, PRACTICE, OR POLICY

⇒ This study opens the future potential of robotic integration in neurovascular intervention and is a stepping stone to a future of remote technology in the field of healthcare.

INTRODUCTION

Cerebral aneurysms affect 2–4% of the population and they can present with a life-threatening rupture with an estimated fatality rate of between 26% and 36%.^{1–3} Once an unruptured aneurysm is detected, the decision to intervene balances the risk–benefit profile of the available interventions against watchful waiting.^{4–9} Treatment options include surgical clipping or endovascular therapies, including coiling, stent-assisted coiling, flow diverters, and intrasaccular devices.^{10 11} The pace of adoption of percutaneous neuroendovascular techniques continues to accelerate, necessitating new technological advancements to improve procedure precision and safety.

The CorPath GRX system is the first robotic platform designed to accommodate the micro guidewires, microcatheters, and microscale movements specific to successful neurovascular interventions. The advent of this robot for percutaneous coronary interventions and peripheral vascular interventions has already



© Author(s) (or their employer(s)) 2023. Re-use permitted under CC BY-NC. No commercial re-use. See rights and permissions. Published by BMJ.

To cite: Mendes Pereira V, Rice H, De Villiers L, et al. *J NeuroIntervent Surg* Epub ahead of print: [please include Day Month Year]. doi:10.1136/jnis-2023-020161

begun to address some of the limitations of endovascular interventions.^{12–20} The CorPath GRX system translates the manual movements of the interventionalist into precision micromovements during navigation and facilitates precision measurement of anatomy to determine lesion length.^{21–22} It also allows the procedure to be performed from a remote, radiation-shielded workstation, which might help to reduce the interventionalist's exposure to ionizing radiation and orthopedic strain.^{17–23–25}

In an initial evaluation of the system in *in vitro* animal neurovascular models,²⁶ Corindus recently implemented software and engineering modifications to its CorPath GRX Robotic System to address neuroendovascular-specific needs and indications: (1) active device fixation, which maintains the placement of both the guidewire and device while the catheter is advanced or retracted by advancing or retracting the guidewire and device to offset the motion of the catheter, (2) limited speed software to reduce the linear movement of the guidewire or device by half, capping its maximum speed at a rate of 6 mm/s, (3) physical modifications to the hardware to securely, and reliably accommodate, the smaller gauge devices common to neurovascular procedures, and (4) software modifications to allow for increased working length to enable target access for microcatheters and an updated user interface to accommodate the new workflow and automated movements. Robotic neurointervention procedures have been performed²⁷ and early feasibility studies have demonstrated successful procedural execution and manual comparable long term follow-ups in complex aneurysm treatments.^{28–31}

This clinical study is a prospective, single-arm, international, multicenter, non-inferiority study to evaluate the effectiveness and safety of the CorPath GRX Robotic System for endovascular cerebral aneurysm embolization compared with historical manual cerebral aneurysm treatment equivalent controls. The primary effectiveness endpoint was defined as successful completion of the robotic-assisted endovascular procedure in the absence of any unplanned conversion to manual for guidewire or microcatheter navigation, embolization coil(s) or intracranial stent(s) deployment, or an inability to navigate vessel anatomy. The primary safety endpoint consisted of a composite of intraprocedural and periprocedural events, including target aneurysmal rupture, vessel perforation or dissection, and major stroke within 24 hours of postprocedure or discharge, whichever occurs first. The secondary endpoints consisted of clinical outcome using the modified Rankin Scale as well as procedure characteristics, including overall robotic and fluoroscopy time, patient radiation exposure, and angiographic assessment of aneurysm occlusion grade according to the Raymond-Roy Classification Scale.

MATERIALS AND METHODS

Study population

The inclusion criteria for this trial were (1) age ≥ 18 years, (2) at least one cerebral aneurysm (unruptured) with indication for endovascular treatment; dome to neck ratio > 1.5 or aneurysm neck width > 4.0 mm, (3) the investigator deemed the procedure appropriate for both manual or robotic-assisted endovascular treatment, (4) patients were informed of the nature of the study and have provided written informed consent. Prospective consecutive recruitment was encouraged, and subjects were excluded if there was (1) failure/unwillingness of the subject to provide informed consent

or if the ethics committee has waived informed consent, (2) the investigator determines that the subject or the neurovascular anatomy is not suitable for robotic-assisted endovascular treatment, (3) women who are pregnant, or (4) people under guardianship or curatorship. Subjects were screened for eligibility based on standard of care assessments, by Oculus Imaging LLC (Knoxville, Tennessee, USA), an independent eligibility committee and administration services provider.

This study was conducted at 10 sites in six countries, including a premarket study in Canada and postmarket study in Australia, Austria, France, Spain, and Switzerland. The study was conducted according to Good Clinical Practice guidelines. Informed consent was obtained from all study participants. Each participating site obtained approval from independent ethics committees and approval as defined in the World Medical Association Declaration of Helsinki, Competent Authority review and approval (if applicable), ISO 14155:2020(E), and the Medical Device Regulations (2017/745) of April 5, 2017. Avania (Bilthoven Netherlands), a contract research organization was responsible for implementing and maintaining monitoring services, with written standard operating procedures to ensure that the clinical investigation was conducted, and data were generated, documented (recorded), and reported in compliance with the clinical investigation plan and above-mentioned guidelines. A surveillance committee, composed of three expert physicians in the field of interventional neuroradiology and/or neurosurgery, who were independent and not directly involved in the conduct of the trial, reviewed and adjudicated all adverse events over the course of the study. Any missing data were provided on request by the independent committee.

This study was reported using the Strengthening the Reporting of Observational Studies in Epidemiology (STROBE) cohort reporting guidelines.³²

Statistical methods

The hypothesis of the endpoints was determined by historical rates from a literature overview (table 1) and pooled results from both arms from three randomized control trials (table 2).

The effectiveness hypothesis was:

- ▶ H_0 : $p \leq 80\%$ (90%–10% non-inferiority margin)
- ▶ H_a : $p > 80\%$ where pP = proportion of subjects with a successfully completed endovascular procedure.

The safety hypothesis was:

- ▶ H_0 : $p \geq 15\%$ (5%+10% non-inferiority margin)
- ▶ H_a : $p < 15\%$ where p = proportion of subjects with a safety event.

Both these endpoints will be summarized along with a two-sided Clopper-Pearson Exact 95% CI. The overall trial sample size was driven by at least 108 subjects enrolled. The secondary endpoints were analyzed with descriptive statistics along with a 95% CI.

The results compiled by this report, include the primary effectiveness and safety endpoints data, monitored through 24 hours postprocedure or hospital discharge, whichever occurred first. All subjects finished this initial timeline for analysis, and the study population is currently being followed up through 180 days for completion of this study.

Table 1 Rates of safety from the literature for endovascular treatment of unruptured and ruptured aneurysms

	N Aneurysms (subjects)	Mortality at discharge	Overall complication rate	Rupture/ perforation rate (%)	Embolism rate (%)
<i>Ruptured intracranial aneurysms</i>					
Algra <i>et al</i> 2019* ³³	73 066 (71 819)	0.3% (0.2–0.4%)	4.96% (4.00, 6.12%)	0.9% (212/18 520) 95% CI (0.6% to 1.3%)	2.82% (437/16 000) 95% CI (2.3% to 3.5%)
Advanced endovascular methods ^{†33}	2248	0.4% (0.2–1.1%)	6.1% (4.3, 8.7%)		
Kawbata <i>et al</i> 2018 ³⁴	1406 (1375)			1.4% (20/1406)	
Coil only	340			1.8% (6/340)	
Stent-assisted coiling	468			0.9% (4/468)	
Balloon-assisted coiling	598			1.7% (10/598)	
From lit review, table 6 in Kawbata ³⁴	7785	0%		1.4% (108/7785)	
Zheng <i>et al</i> 2016 ³⁵	1127			1% (11/1127)	
Santillan <i>et al</i> 2013 ³⁶	217	0%		1.1% (3/217)	
Shigematsu <i>et al</i> 2013 ³⁷	4767			1.4% (65/4767)	
Oishi <i>et al</i> 2012 ³⁸	500	0%		1.4% (7/500)	
Im <i>et al</i> 2009 ³⁹	435	0%		0.9% (4/435)	
Pierot <i>et al</i> 2008 ⁴⁰	739 (649)	0.3%		2.4% (18/739)	7.3% (29/398)
(ATENA study)	(700 procs)	1.4% (1 month)	15.4%	2.6% (18/700)	7.1% (per proc)
Pierot <i>et al</i> 2009 ⁴¹ (ATENA study, coil alone results)	325	0.9% (3/325)	10.8% (35/325)	2.1% (7/325)	6.2% (20/325)
<i>Ruptured intracranial aneurysms</i>					
Zhang <i>et al</i> 2019 ⁴²	1004	All cause 9.5% (5.8%, 13.2%) procedure related 1.8% (0.9, 2.7%)	22.7%‡ 95% CI (15.1% to 30.3%)		
Cognard <i>et al</i> 2011 ⁴³ (CLARITY GDC study)	405	1.50%		3.7% (15/405)	13.3% (54/405)
Pierot <i>et al</i> 2011 (CLARITY) ⁴⁴	608	5.1% (31/608) cum. treatment-related morbidity/mortality rate 19.6% (119/608) with morbidity/mortality rate	17.4% (106/608)	4.6% (28/608)	

*Rates presented are pooled crude risk from meta-analysis modeling.
 †Advanced=stent-assisted coiling, balloon-assisted coiling, flow diverting stents or Woven Endobridge devices.
 ‡Definition not specified.

RESULTS

The study is summarized in figure 1. A total of 130 subjects were screened and provided written informed consent between August 2020 and April 2022. The complete follow-up period is 180 days. Nine of these subjects were

screen failures and four device malfunction during set-up, leading to 117 subjects being included in the analysis population, completed by 14 operators.

Four device malfunctions occurred during procedural setup: two extended-reach arm failures, one software

Table 2 Pooled results from three randomized controlled trials

HELPS Trial	N	Mortality	Procedural aneurysm rupture	Thromboembolic complication	Technical success		
White <i>et al</i> 2008 ⁴⁵	499	2.2% (11/499) At discharge	3.4% (17/499)	10.2% (51/499)	96.6% (482/499)		
MAPS Trial	N	Mortality	Overall complication rate	Clinical event committee adjudicated	Technical success		
				Rupture/ re-rupture	Ischemic stroke	Hemorrhagic stroke	
McDougall <i>et al</i> 2014 ⁴⁶	626	0.2% (1/624) Periproc 2.4% (15/626) 30 day	14.9% (93/624)	0.3% (2/626)	3.4% (21/626)	1.1% (7/626)	97.1% (608/626)
Cerecyte Trial	N	Mortality	Overall complication rate	Procedural aneurysm rupture	Thromboembolic complication	Neuro deterioration	Technical success
Coley <i>et al</i> 2012 ⁴⁷	497	0% within 24 hours 0.4% (2/497) discharge	12.3% (61/497)	3.6% (18/497)	5.6% (28/497)	3.0% (15/497)	97.2% (483/497)




Figure 1 A flowchart of the executive trial summary.

initiation failure, and one cassette and/or console error. As the procedure was not performed robotically, these subjects were not included in the analysis cohort. Three of these were encountered at the same site. Both extended-reach arm malfunctions happened consecutively, and inspection revealed the cables were not securely connected, with the issue being resolved immediately once the cables were securely attached. The initiation failure happened due to the robot power supply being connected to an outlet controlled by the fluoroscopy pedal, leading to the robotic power being turned off when the fluoroscopy pedal was pressed, which led the robot to not start while the other equipment was in use and was resolved on dealing with this concern. The last one was identified as a cassette and/or console error, which was encountered after connecting the devices, on which an error was noted and prevented the continuation of the procedure robotically.

Table 3 summarizes the subject demographics and medical history. The mean age of subjects was 56.6 ± 12.7 years, with 74% of the subjects being female. History of smoking was prevalent in 59% of subjects. Hypertension in 49.6% of subjects was the leading comorbidity reported, followed by multiple aneurysms in 38.5% and dyslipidemia in 22.2% of subjects.

Aneurysm characteristics are summarized in table 4. Most of the aneurysms were located at the internal carotid artery (27.9%), followed by the anterior communicating artery (22.1%) and middle cerebral artery (9.8%). The mean aneurysm height and neck width was 5.7 ± 2.6 mm and 3.5 ± 1.4 mm, respectively. The mean aneurysm dome to neck ratio was 1.7 ± 0.4 .

Procedural characteristics are presented in table 5. The access site was femoral in 60.7% and radial in 37.6% of subjects. All 117 subjects were treated with the primary coiling procedure, with 58 subjects requiring adjunct stenting. The mean number of coils implanted was 4.8 ± 3.9 . Of the 58 stenting subjects, 50 received one stent, alongside eight subjects receiving two or more stents. The overall

procedure time was 117.3 ± 47.3 min, with the robotic procedure time of 59.4 ± 32.6 min. The total fluoroscopy time was 51.9 ± 27.4 min with total contrast of 185.4 ± 82.3 cc.

Primary effectiveness success was achieved with the procedure completed robotically in 94.0% (110/117) (88.1%, 97.6%) subjects. Unplanned manual conversion was done in 6.0% (7/117) subjects. Two device malfunctions occurred: (1) the control console malfunction was attributed to the joystick capacitive functionality, which was

Table 3 Baseline demographics and clinical characteristics

Baseline demographics and clinical characteristics	Subjects (n=117)
Age, mean (range)	56.6±12.7 (22 to 79)
Sex	Female 74.4%, males 25.6%
BMI (n=115)	26.7±5.5 (15.6 to -44.3)
Medical history	
History of smoking	59.0% (69/117)
Hypertension	49.6% (58/117)
Multiple aneurysms	38.5% (45/117)
Dyslipidemia	22.2% (26/117)
Family history of UIA or SAH	17.9% (21/117)
History of SAH	17.1% (20/117)
Diagnosed with other rare disease/condition	10.3% (12/117)
Diabetes	8.5% (10/117)
Possible excess alcohol use	6.0% (7/116)
Indications	
Asymptomatic or incidental aneurysm	54.7% (64/117)
Headache	34.2% (40/117)
Aneurysm(s) with a prior SAH from a separate aneurysm	6.0% (7/117)
Stroke	1.7% (2/117)
BMI, body mass index; SAH, subarachnoid hemorrhage; UIA, unruptured intracranial aneurysm.	

Table 4 Aneurysm characteristics

Aneurysm characteristics	n=122
Location	
Internal carotid artery	27.9% (34/122)
Anterior communicating artery	22.1% (27/122)
Middle cerebral artery	9.8% (12/122)
Posterior communicating artery	9.0% (11/122)
Basilar artery	8.2% (10/122)
Anterior cerebral artery	3.3% (4/122)
Anterior choroidal artery	2.5% (3/122)
Posterior cerebral artery	1.6% (2/122)
Other	15.6% (19/122)
Aneurysm height (mm)	5.7±2.6 (1.5 to 15)
Aneurysm neck width (mm)	3.5±1.4 (1.2 to 9.2)
Dome to neck ratio	1.7±0.4 (0.8 to 3.0)
Relationship to parent wall	
Bifurcation	41.0% (50/122)
Sidewall	59.0% (72/122)

not actuating when touched; an investigation identified a loosened joystick cover mounting screws impact from shipping or improper handling, and (2) the cassette malfunction was attributed to an unknown root cause. Five procedural limitations were attributed as follows: three, loss of working length attributed to intermediary catheter placement that contributed to insufficient working length for the microcatheter; one due to off-label microcatheter use during procedure; and one attributed to early detachment of the second coil not allowing its insertion into the aneurysm. **Table 6** summarizes the primary endpoint results.

Primary safety endpoint occurred in 3.4% (4/117) (0.9%, 8.5%) subjects, with two major strokes and two intraprocedural aneurysm ruptures. A total of 24 adjudicated adverse events were reported. Stroke was the most common serious

Table 5 Procedural characteristics

Procedural characteristics	
Access Site	
Femoral	60.7% (71/117)
Radial	37.6% (44/117)
Brachial	1.7% (2/117)
No. of coils implanted (n=117)	4.8±3.9 (0 to 31)
Stents implanted	
0	50.4% (59/117)
1	42.7% (50/117)
≥2	6.9% (8/117)
Maximum stent diameter (mm) (n=58)	3.9±0.7 (3 to 6)
Overall procedure time (min) (n=115)	117.3±47.3 (17 to 259)
Robotic procedure time (min) (n=116)	59.4±32.6
Fluoroscopy time (min) (n=117)	51.9±27.4 (10 to 157)
Total contrast (cc) (n=116)	185.4±82.3 (60 to 530)
Subject radiation exposure (mGy.cm ²) (n=104)	6 886 135.5±25 767 850.8 (16 814.7 to 143 730 000)
Air Kerma (mGy) (n=98)	2220.4±1526.9 (1914 to 2527)

Table 6 Primary endpoints

Endpoint	Analysis cohort (Subjects)
Primary Effectiveness	94.0% (110/117) (88.1%, 97.6%)
Procedure Completed Robotically	94.0% (110/117)
Unplanned Manual Conversion	6.0% (7/117)
Device Malfunction	1.7% (2/117)
Control Console Failure	0.9% (1/117)
Cassette Failure	0.9% (1/117)
Procedural Limitation	4.3% (5/117)
Loss of Working Length	2.6% (3/117)
Use of Off-label Microcatheter	0.9% (1/117)
Coil Deployment	0.9% (1/117)
Primary Safety	3.4% (4/117) (0.9%, 8.5%)
Stroke (Major)	1.7% (2/117)
Intraprocedural Aneurysm Rupture	1.7% (2/117)

adverse event reported in 12 subjects, with 10 experiencing a minor stroke and only two with a major stroke. Vascular access site complications were reported in four subjects and intraprocedural aneurysm ruptures in two. Electrolyte disorder occurred in two subjects; and hydrocephalus, vasospasm, respiratory failure, and urine infection were each reported once. **Table 7** summarizes these adverse events, and **table 8** details the presentation and clinical outcome of each adverse event.

A postprocedure Raymond-Roy Classification Scale score of 1 was achieved in 64.5% of subjects, signifying complete obliteration; followed by 4.5% having residual neck and the remaining 31% with a residual aneurysm. The modified Rankin Scale (mRS) score was 0 for 78.2% of the subjects post procedure, with no subjects having mRS score ≥3. **Tables 9 and 10** summarize this information. Both, the Raymond-Roy Classification Scale and mRS will be continued to follow-up for all subjects till 180 days, as part of the trial design.

DISCUSSION

The primary effectiveness success achieved in 94% (110/117) subjects was slightly lower than technical success of the three randomized controlled trials: HELPS, MAPS, and Cerecyte Trial, ranging from 96.6 to 97.2%. Of the seven unplanned conversions to manual treatment that were attributed to not meeting the primary effectiveness endpoint, only two were attributed to device malfunctions that happened during the

Table 7 All adverse events within 24 hours (adjudicated)

Adverse event type	No. of subjects (%)
Stroke (minor)	10/117 (8.5%)
Stroke (major)	2/117 (1.7%)
Intraprocedural aneurysm rupture	2/117 (1.7%)
Vascular access site complication	4/117 (3.4%)
Electrolyte disorder	2/117 (1.7%)
Hydrocephalus	1/117 (0.9%)
Vasospasm	1/117 (0.9%)
Respiratory failure	1/117 (0.9%)
Urine infection	1/117 (0.9%)

Table 8 Adverse event presentation and outcome

All adverse events: classification and outcome	
Stroke (Minor)	
01–29	Seizure episode - home discharge on third postoperative day with mRS score 1
01–35	Headache and loss of strength/sensation episode overnight - home discharge on third postoperative day with mRS score 0
02–19	NIHSS score 2 Stroke - home discharge on fifth postoperative day with mRS score 1
02–20	NIHSS score 1 Stroke - private rehab discharge on ninth postoperative day with mRS score 1
02–22	NIHSS score 2 Stroke - home discharge on fifth postoperative day with mRS score 1
02–23	Left extremities' weakness - home discharge on sixth postoperative day with mRS score 0
02–25	Right upper limb weakness and sluggish pupil - home discharge on ninth postoperative day with mRS score 0
02–27	Aphasia/altered sensation/panic episode - home discharge on fifth postoperative day with mRS score 1
07–10	Somnolent with acute respiratory insufficiency – home discharge on ninth postoperative day with mRS score 1
11–02	'Discrete micro-ischemia' on Imaging – home discharge on second postoperative day with mRS score 0
Stroke (major)	
01–05	NIHSS 1 Stroke with anterograde amnesia – home discharge on second postoperative day with mRS score 1
04–06	Right hemi-paresis requiring intra-arterial fibrinolysis - home discharge on third postoperative day with mRS score 1
Intraprocedural aneurysm rupture	
01–19	Rupture during second coil deployment, with hemostasis rapidly achieved with heparin reversion and further coiling
04–04	Rupture during coil deployment, with extravasation stopped via manual carotid compression and remodeling balloon inflation at aneurysmal neck
Vascular access site complication	
01–33	Minor hematoma at puncture site
05–10	Pseudoaneurysm requiring 2 units of blood transfusion
05–19	Hematoma requiring a follow-up CT scan, leading to extended hospitalization
10–11	Retroperitoneal hematoma requiring 1 unit of blood transfusion
Electrolyte disorder	
04–04	During hospital course of same subject with intraprocedural aneurysm rupture
04–06	During hospital course of same subject with major stroke
Hydrocephalus	
04–04	Same subject with intraprocedural aneurysm rupture, requiring external ventricular drainage placement
Vasospasm	
04–06	Same subject with major stroke - focal vasospasm, responded to nimodipine infusion
Respiratory failure	
07–10	Same subject with minor stroke – acute respiratory insufficiency due to chronic emphysema
Urine Infection	
05–08	Antibiotic treatment requiring prolonged hospitalization

mRS, modified Rankin Scale; NIHSS, National Institutes of Health Stroke Scale.

procedure, with the remaining five procedural limitations. [Table 11](#) summarizes the literature comparison.

In subjects treated with the CorPath GRX System, a total of 25 adverse events subjects were reported. All these events were reviewed and classified by an independent clinical events committee. Ten of these events were attributed to minor strokes with a National Institutes of Health Stroke Scale score <3, and each of them required no further

intervention, with the subjects being clinically observed and discharged with an mRS score of either 0 or 1. Of the five main serious events that constituted primary safety endpoint failures, two were major strokes, and both subjects were discharged home clinically stable on the second and third postoperative, day respectively, with an mRS score of 1. Previous literature highlights thromboembolic

Table 9 Raymond-Roy Classification scale

Raymond-Roy Classification Scale (postprocedure) n=110	
1	64.5% (71/110)
2	4.5% (5/110)
3a	25.5% (28/110)
3b	5.5% (6/110)

Table 10 Modified Rankin Scale

Clinical outcome		
Modified Rankin Scale	Preprocedure	Postprocedure
n=mean±SD (range)	115=0.4±0.6 (0 to 2)	78=0.3±0.5 (0 to 2)
0	65.2% (75/115)	78.2% (61/78)
1	28.7% (33/115)	17.9% (14/78)
2	6.1% (7/115)	3.8% (3/78)

Table 11 Aneurysm treatment prospective studies literature used for comparison

	HELPS Trial	MAPS Trial	Cerecyte Trial	CorPath GRX
Technical success	96.6 %	97.1 %	97.2 %	94.0 %
Intraprocedural aneurysm rupture	3.4 %	2.2 %	3.6 %	2.6 %
Thromboembolic complications	10.2 %	4.5 %	5.6 %	10.25 %
Mortality	2.2% at discharge	0.2% periprocedural	0.4% before discharge	0% at discharge
Raymond-Roy Scale postprocedure	Operator self-assessed	Core-laboratory review	Core-laboratory review	Core-laboratory review
1	47.5 %	36.1 %	26.2 %	64.5%
2	34.5 %	26.2 %	49.5 %	4.5 %
3	18.0 %	37.7 %	24.3 %	31.0 %
Clinical assessment at discharge	WFNS		mRS	mRS
	0	38.5 %	0	81.8 %
	1–2	55.3 %	1	15.9 %
	≥ 3	6.0 %	2	2.3 %
			≥ 3	0 %

mRS, modified Rankin Scale; WFNS, World Federation of Neurological Surgeons.

complications ranging from 5.6% in the Cerecyte Trial to 10.2% in the HELPS Trial. Combining both minor and major stroke, constitute 10.26% (12/117), which is comparable to the upper limit reported by these randomized controlled trials described, all of our strokes concluded with the subjects discharged clinically stable with an mRS score ≤ 1 . The two remaining events considered primarily safety endpoint failures, were intraprocedural aneurysm ruptures (1.7% subjects), which occurred within the range of these trials mentioned above: 2.2% in MAPS to 3.6% in Cerecyte.

Comparison of the core-laboratory reviewed Raymond-Roy Scale postprocedure showed that CorPath GRX had the highest cohort of complete obliteration score in 64.5% subjects, and residual aneurysm in 31.0% of subjects, which was in comparison with the range of 24.3% in Cerecyte to 37.7% in MAPS. The modified Rankin Scale was almost similar to the those reported by Cerecyte, with no patient having an mRS score of ≥ 3 .

Limitations

The current generation of robots has technical limitations that should be taken into consideration when planning surgery: the ability to manage a total of only one microcatheter and one microwire or device at a time; the requirement for the access system be placed manually; and the length of working time for a procedure. Future generations of the robotic system are expected to overcome these issues.

CONCLUSION

This first-of-its-kind robotic-assisted neurovascular trial demonstrates the effectiveness and safety of the CorPath GRX System for endovascular embolization cerebral aneurysm procedures.

Author affiliations

¹Division of Neurosurgery, Department of Surgery, St Michael's Hospital, University of Toronto, Toronto, Ontario, Canada

²Department of Neurointervention, Gold Coast University Hospital, Southport, Queensland, Australia

³Department of Interventional Neuroradiology, Hopital Universitaire Pitie Salpetriere, Paris, France

⁴Department of Neurointervention, Hospital Vall d'Hebron, Barcelona, Catalunya, Spain

⁵Department of Medical Imaging, Toronto Western Hospital, Toronto, Ontario, Canada

⁶Department of Neuroradiology, Centre Hospitalier Universitaire de Montpellier,

Montpellier, France

⁷Department of Diagnostic and Interventional Neuroradiology, Inselspital Universitätsklinik Bern, Bern, Switzerland

⁸Department of Interventional Neuroradiology and Endovascular Neurosurgery, Hospital Clinico Universitario de Valladolid, Valladolid, Spain

⁹Department of Neurosurgery, Paracelsus Medical University, Salzburg, Austria

¹⁰Division of Neurosurgery, Prisma Health, Greenville, South Carolina, USA

¹¹Department of Interventional Neuroradiology, Fondation Rothschild, Paris, France

Twitter Vitor Mendes Pereira @VitorMendesPer1, Nicole M Cancelliere @NMCancelliere and Mario Martínez-Galdámez @Doctorgaldamez

Contributors Guarantor: VMP. Conception and design: VMP, MP and RDT. Acquisition of data: all authors. Analysis and interpretation of data: VMP, MP and RDT. Drafting the article: VMP, MP, RB, RDT, HR and XYEL. Critically revising the article: all authors. Approved the final version of the manuscript: all authors.

Funding This study was funded by Siemens Healthineers Endovascular Robotics.

Competing interests VMP is an unpaid consultant for Siemens Healthineers Endovascular Robotics. The following authors have one or more competing interests: HR, NS, FC, DH, MP, JG, MK-O, RDT, RB. These competing interests include board membership, consultancy, employment, expert testimony grants, contract research, lectures/other education events, speakers' bureau, and receipt of equipment or supplies, royalties, stock/stock options/other forms of ownership.

Patient consent for publication Consent obtained directly from patient(s)

Ethics approval This study involves human participants and was approved by The Unity Health Toronto Research Ethics Board; REB 20-333; Les Comités de Protection des Personnes IORG0009918; Gold Coast Hospital and Health Service Human Research Ethics Committee EC00160; University Health Network Research Ethics Board 19-5517; Research Ethics Committee with Medicines of the Hospital Universitari Vall d'Hebron IRB00002850; Gesundheits-, Sozial- und Integrationsdirektion Kantonale Ethikkommission für die Forschung, Ethikkommission Bern 2021-D0024; Ethikkommission für das Bundesland Salzburg 1222/2021; Comité de Ética de la Investigación con Medicamentos Área de Salud Valladolid Este Hospital Clínico Valladolid CASVE-PS-20-443. Participants gave informed consent to participate in the study before taking part.

Provenance and peer review Not commissioned; externally peer reviewed.

Data availability statement Data are available upon reasonable request.

Open access This is an open access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited, appropriate credit is given, any changes made indicated, and the use is non-commercial. See: <http://creativecommons.org/licenses/by-nc/4.0/>.

ORCID iDs

Vitor Mendes Pereira <http://orcid.org/0000-0002-6804-3985>

Alejandro Tomasello <http://orcid.org/0000-0002-6736-1535>

Nicole M Cancelliere <http://orcid.org/0000-0002-8703-4304>

Gregory Gasco <http://orcid.org/0000-0002-0817-7324>

Mario Martínez-Galdámez <http://orcid.org/0000-0002-8024-4712>
 Monika Killer-Oberpfalzer <http://orcid.org/0000-0001-5572-7694>

REFERENCES

- 1 Lawton MT, Vates GE. Subarachnoid hemorrhage. *N Engl J Med* 2017;377:257–66.
- 2 Lantigua H, Ortega-Gutierrez S, Schmidt JM, et al. Subarachnoid hemorrhage: who dies, and why *Crit Care* 2015;19:309.
- 3 Connolly ES Jr, Rabinstein AA, Carhuapoma JR, et al. Guidelines for the management of aneurysmal subarachnoid hemorrhage. *Stroke* 2012;43:1711–37. 10.1161/STR.0b013e3182587839
- 4 Nasr DM, Brown RD. Management of unruptured intracranial aneurysms. *Curr Cardiol Rep* 2016;18:86.
- 5 Ajiboye N, Chalouhi N, Starke RM, et al. Unruptured cerebral aneurysms: evaluation and management. *ScientificWorldJournal* 2015;2015:954954.
- 6 Pierot L, Gawlitza M, Soize S. Unruptured intracranial aneurysms: management strategy and current endovascular treatment options. *Expert Rev Neurother* 2017;17:977–86.
- 7 Boulouis G, Rodriguez-Régent C, Rasolonjatovo EC, et al. Unruptured intracranial aneurysms: an updated review of current concepts for risk factors, detection and management. *Rev Neurol (Paris)* 2017;173:542–51.
- 8 Flemming KD, Lanzino G. Management of Unruptured intracranial aneurysms and cerebrovascular malformations. *Continuum (Minneapolis)* 2017;23:181–210.
- 9 Williams LN, Brown RD. Management of unruptured intracranial aneurysms. *Neuro Clin Pract* 2013;3:99–108.
- 10 Darsaut TE, Findlay JM, Magro E, et al. Surgical clipping or endovascular coiling for unruptured intracranial aneurysms: a pragmatic randomised trial. *J Neurol Neurosurg Psychiatry* 2017;88:663–8.
- 11 Ruan C, Long H, Sun H, et al. Endovascular coiling vs. surgical clipping for unruptured intracranial aneurysm: a meta-analysis. *Br J Neurosurg* 2015;29:485–92.
- 12 Smitson CC, Ang L, Pourdjabbar A, et al. Safety and feasibility of a novel, second-generation Robotic-assisted system for percutaneous coronary intervention: first-in-human report. *J Invasive Cardiol* 2018;30:152–6.
- 13 Almasoud A, Walters D, Mahmud E. Robotically performed excimer laser coronary atherectomy: proof of feasibility. *Catheter Cardiovasc Interv* 2018;92:713–6.
- 14 Swaminathan RV, Rao SV. Robotic-assisted Transradial diagnostic coronary angiography. *Catheter Cardiovasc Interv* 2018;92:54–7.
- 15 Mahmud E, Dominguez A, Bahadorani J. First-In-human robotic percutaneous coronary intervention for unprotected left main stenosis. *Cathet. Cardiovasc. Intervent.* 2016;88:565–70.
- 16 Behnamfar O, Pourdjabbar A, Yalvac E. First case of robotic percutaneous vascular intervention for below-the-knee peripheral arterial disease. *J Invasive Cardiol* 2016;28:E128–31.
- 17 Mahmud E, Schmid F, Kalmar P, et al. Feasibility and safety of robotic peripheral vascular interventions: results of the RAPID Trial. *JACC Cardiovasc Interv* 2016;9:2058–64.
- 18 Weisz G, Smilowitz NR, Metzger DC, et al. The association between experience and proficiency with robotic-enhanced coronary intervention—insights from the PRECISE multi-center study. *Acute Card Care* 2014;16:37–40.
- 19 Smilowitz NR, Moses JW, Sosa FA. Robotic-enhanced PCI compared to the traditional manual approach. *J Invasive Cardiol* 2014;26:318–21.
- 20 Weisz G, Metzger DC, Caputo RP, et al. Safety and feasibility of robotic percutaneous coronary intervention. *J Am Coll Cardiol* 2013;61:1596–600.
- 21 Campbell PT, Kruse KR, Kroll CR, et al. The impact of precise robotic lesion length measurement on stent length selection: ramifications for stent savings. *Cardiovasc Revasc Med* 2015;16:348–50.
- 22 Bergman P, Blacker SJ, Kottenstette N, et al. Robotic-assisted percutaneous coronary intervention. In: Abedin-Nasab MH, ed. *Handbook of Robotic and Image-Guided Surgery*. Elsevier, 2020: 341–62.
- 23 Andreassi MG, Piccaluga E, Gargani L, et al. Subclinical carotid atherosclerosis and early vascular aging from long-term low-dose ionizing radiation exposure. *JACC: Cardiovasc Interv* 2015;8:616–27.
- 24 Andreassi MG, Piccaluga E, Guagliumi G, et al. Occupational health risks in cardiac catheterization laboratory workers. *Circ Cardiovasc Interv* 2016;9:e003273.
- 25 Venneri L, Rossi F, Botto N, et al. Cancer risk from professional exposure in staff working in cardiac catheterization laboratory: Insights from the National Research Council's Biological Effects of Ionizing Radiation VII Report. *Am Heart J* 2009;157:118–24.
- 26 Britz GW, Tomas J, Lumsden A. Feasibility of robotic-assisted neurovascular interventions: initial experience in flow model and porcine model. *Neurosurgery* 2020;86:309–14.
- 27 Mendes Pereira V, Cancelliere NM, Nicholson P, et al. First-In-human, robotic-assisted neuroendovascular intervention. *J Neurointerv Surg* 2020;12:338–40.
- 28 Mendes Pereira V, Nicholson P, Cancelliere NM, et al. Feasibility of robot-assisted neuroendovascular procedures. *J Neurosurg* 2021;136:992–1004.
- 29 Cancelliere NM, Lynch J, Nicholson P, et al. Robotic-assisted intracranial aneurysm treatment: 1 year follow-up imaging and clinical outcomes. *J Neurointerv Surg* 2022;14:1229–33.
- 30 Weinberg JH, Sweid A, Sajja K, et al. Comparison of robotic-assisted carotid stenting and manual carotid stenting through the transradial approach. *J Neurosurg* 2020;135:21–8.
- 31 Beaman C, Saber H, Tateshima S. A technical guide to robotic catheter angiography with the corindus corpath GRX system. *J Neurointerv Surg* 2022;14:1284.
- 32 von Elm E, Altman DG, Egger M, et al. The Strengthening the Reporting of Observational Studies in Epidemiology (STROBE) statement: guidelines for reporting observational studies. *Epidemiology* 2007;18:800–4.
- 33 Algra AM, Lindgren A, Vergouwen MDI, et al. Procedural clinical complications, case-fatality risks, and risk factors in endovascular and neurosurgical treatment of unruptured intracranial aneurysms: a systematic review and meta-analysis. *JAMA Neurol* 2019;76:282–93.
- 34 Kawabata S, Imamura H, Adachi H, et al. Risk factors for and outcomes of Intraprocedural rupture during endovascular treatment of unruptured intracranial aneurysms. *J Neurointerv Surg* 2018;10:362–6.
- 35 Zheng Y, Liu Y, Leng B, et al. Perioperative complications associated with Endovascular treatment of intracranial aneurysms in 1764 cases. *J Neurointerv Surg* 2016;8:152–7.
- 36 Santillan A, Gobin YP, Mazura JC, et al. Balloon-assisted coil embolization of intracranial aneurysms is not associated with increased perioperative complications. *J Neurointerv Surg* 2013;5 Suppl 3:iii56–61.
- 37 Shigematsu T, Fujinaka T, Yoshimine T, et al. Endovascular therapy for asymptomatic unruptured intracranial aneurysms. *Stroke* 2013;44:2735–42.
- 38 Oishi H, Yamamoto M, Shimizu T, et al. Endovascular therapy of 500 small asymptomatic unruptured intracranial aneurysms. *AJNR Am J Neuroradiol* 2012;33:958–64.
- 39 Im S-H, Han MH, Kwon O-K, et al. Endovascular coil embolization of 435 small asymptomatic unruptured intracranial aneurysms: procedural morbidity and patient outcome. *AJNR Am J Neuroradiol* 2009;30:79–84.
- 40 Pierot L, Spelle L, Vitry F, et al. Immediate clinical outcome of subjects harboring unruptured intracranial aneurysms treated by endovascular approach: results of the ATENA study. *Stroke* 2008;39:2497–504.
- 41 Pierot L, Spelle L, Leclerc X, et al. Endovascular treatment of unruptured intracranial aneurysms: comparison of safety of remodeling technique and standard treatment with coils. *Radiology* 2009;251:846–55.
- 42 Zhang X, Zuo Q, Tang H, et al. Stent assisted coiling versus non-stent assisted coiling for the management of ruptured intracranial aneurysms: a meta-analysis and systematic review. *J Neurointerv Surg* 2019;11:489–96.
- 43 Cognard C, Pierot L, Anxionnat R, et al. Results of embolization used as the first treatment choice in a consecutive nonselected population of ruptured aneurysms: clinical results of the clarity GDC study. *Neurosurgery* 2011;69:837–41.
- 44 Pierot L, Cognard C, Anxionnat R, et al. Remodeling technique for endovascular treatment of ruptured intracranial aneurysms had a higher rate of adequate postoperative occlusion than did conventional coil embolization with comparable safety. *Radiology* 2011;258:546–53.
- 45 White PM, Lewis SC, Nahser H, et al. Hydrocoil endovascular aneurysm occlusion and packing study (HELPS trial): procedural safety and operator-assessed efficacy results. *AJNR Am J Neuroradiol* 2008;29:217–23.
- 46 McDougall CG, Johnston SC, Gholkar A, et al. Bioactive versus bare platinum coils in the treatment of intracranial aneurysms: the MAPS (matrix and platinum science). *AJNR Am J Neuroradiol* 2014;35:935–42.
- 47 Coley S, Sneade M, Clarke A, et al. Cerecyte coil trial: procedural safety and clinical outcomes in subjects with ruptured and unruptured intracranial aneurysms. *AJNR Am J Neuroradiol* 2012;33:474–80.

J NeuroIntervent Surg: first published as 10.1136/jnis-2023-020161 on 4 October 2023. Downloaded from <http://jnis.bmj.com/> on July 18, 2024 by guest. Protected by copyright.